Techline Technologies, Inc.
Device Specifications PTT-B70
Nose in a Box<sup>™</sup>





# **PTT-B70**

## Nose in a Box<sup>TM</sup>

Patent Nos. US10,217,380, US10,535,282, US10,586,470, US11,195,434, US11,195,435, Patents Pending

TECHLINE ® Device Specifications

March 11, 2022

### Description-

The Nasopharyngeal Airway (NPA) trainer, or Nose in a Box<sup>™</sup> is designed to be simple, durable, economical, and portable. This simple device allows the demonstration and practice of NPA insertion. (NPA Device not included). This device also can be used for training the administration of naloxone (NARCAN).

#### Construction-

The representative skin and tissue of the PTT-B70 is constructed of high quality silicone rubber, with layers of coloration that provide a depth and realism to the product. In addition, the SOFTTECH® material has a realistic feel, and is very durable and UV resistant.


#### Operation-

The Nose in a Box<sup>™</sup> can be used to train in nasopharyngeal intubation and Naloxone (Narcan) administration. Practice Narcan administration by inserting the tip of the nozzle into one nostril and pressing the plunger firmly. Lubricate the nasopharyngeal airway with water-soluble lubricant or anesthetic jelly. Insert the airway posteriorly to the nasal cavity.

#### Maintenance-

The device should be rinsed in clean, warm water and allowed to dry before storage. The simulation may be removed from the storage case for cleaning. The device may be washed with mild detergents like dish soap or cleaned with a disinfecting agent like isopropyl alcohol or a diluted bleach solution, as it is durable, and chemically resistant. Allow the device to dry fully before storage to prevent mold or mildew growth.

Techline Technologies, Inc. Device Specifications PTT-B70 Nose in a Box™



PTT-B70 Nose in a Box<sup>™</sup>